

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL121020\
 Data File : VL036132.D
 Acq On : 10 Dec 2020 15:50
 Operator : SY/AP
 Sample : VL1210ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1210ABL01

Manual Integrations
 APPROVED

MMDadoda
 12/15/2020 4:51:17 PM

Quant Time: Dec 11 03:47:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 0
 QLast Update : Mon Dec 07 16:56:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1361373	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.17	114	2801009	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.08	117	2489700	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2121638	10.73	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.86	43	35414m	0.216	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL121020\
Data File : VL036132.D
Acq On : 10 Dec 2020 15:50
Operator : SY/AP
Sample : VL1210ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1210ABL01

Manual Integrations
APPROVED

MMDadoda
12/15/2020 4:51:17 PM

Quant Time: Dec 11 03:47:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL120720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec
QLast Update : Mon Dec 07 16:56:49 2020
Response via : Initial Calibration

