

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL121118\
 Data File : VL032954.D
 Acq On : 11 Dec 2018 14:41
 Operator : SY/AP
 Sample : VL1211ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1211ABL01

Manual Integrations
 APPROVED

MMDadoda
 12/13/2018 7:58:36 AM

Quant Time: Dec 12 16:07:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 0
 QLast Update : Fri Dec 07 09:35:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1213687	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	3006168	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	2938931	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2153200	9.92	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.20%

Target Compounds					Ovalue
15) Acetone	3.94	43	10992m	0.106	ppbv
20) Methylene Chloride	4.42	84	17311m	0.340	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL121118\
Data File : VL032954.D
Acq On : 11 Dec 2018 14:41
Operator : SY/AP
Sample : VL1211ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1211ABL01

Manual Integrations
APPROVED

MMDadoda
12/13/2018 7:58:36 AM

Quant Time: Dec 12 16:07:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec
QLast Update : Fri Dec 07 09:35:17 2018
Response via : Initial Calibration

