

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL121118\
 Data File : VL032974.D
 Acq On : 12 Dec 2018 3:48
 Operator : SY/AP
 Sample : J6262-04DL 400X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL

Manual Integrations
 APPROVED

MMDadoda
 12/13/2018 7:58:54 AM

Quant Time: Dec 12 04:30:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 0
 QLast Update : Fri Dec 07 09:35:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1216466	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2951918	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	2949716	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2029060	9.31	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	93.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.02	51	16664m	0.181	ppbv	
22) trans-1,2-Dichloroethene	4.97	96	14698	0.232	ppbv #	68
31) cis-1,2-Dichloroethene	5.64	61	27927	0.267	ppbv #	84
38) Trichloroethene	7.96	130	141783m	1.569	ppbv	
52) Tetrachloroethene	11.38	164	832753	10.544	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL121118\
Data File : VL032974.D
Acq On : 12 Dec 2018 3:48
Operator : SY/AP
Sample : J6262-04DL 400X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL

Manual Integrations
APPROVED

MMDadoda
12/13/2018 7:58:54 AM

Quant Time: Dec 12 04:30:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec
QLast Update : Fri Dec 07 09:35:17 2018
Response via : Initial Calibration

