

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
 Data File : VL041723.D
 Acq On : 11 Dec 2024 16:17
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 12/13/2024
 Supervised By :Semsettin Yesilyurt 12/13/2024

Quant Time: Dec 12 13:17:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Dec 12 13:14:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	276165	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.972	114	1374483	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	1217003	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	800482	10.005	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.100%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.583	62	1539	0.098	ppbv	97
32) 1,1,1-Trichloroethane	3.379	97	6130	0.095	ppbv	97
35) Carbon Tetrachloride	3.781	117	5606m	0.088	ppbv	
38) Trichloroethene	4.564	130	5196	0.093	ppbv	94
52) Tetrachloroethene	8.247	164	5884	0.103	ppbv	92
54) 1,2-Dibromoethane	7.677	107	5743	0.092	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.979	83	7566	0.095	ppbv	96
79) Naphthalene	13.529	128	9818	0.082	ppbv #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
Data File : VL041723.D
Acq On : 11 Dec 2024 16:17
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/13/2024
Supervised By :Semsettin Yesilyurt 12/13/2024

Quant Time: Dec 12 13:17:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Thu Dec 12 13:14:06 2024
Response via : Initial Calibration

