

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
 Data File : VL041724.D
 Acq On : 11 Dec 2024 16:48
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 12/13/2024
 Supervised By :Semsettin Yesilyurt 12/13/2024

Quant Time: Dec 12 13:17:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Dec 12 13:14:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	271460	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.975	114	1347505	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.901	117	1202636	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	782948	9.903	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.000%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.583	62	583	0.038	ppbv #	64
32) 1,1,1-Trichloroethane	3.379	97	2101m	0.033	ppbv	
38) Trichloroethene	4.567	130	1777m	0.032	ppbv	
52) Tetrachloroethene	8.244	164	1822m	0.032	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.985	83	2549m	0.032	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
Data File : VL041724.D
Acq On : 11 Dec 2024 16:48
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Dec 12 13:17:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Dec 12 13:14:06 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/13/2024
Supervised By :Semsettin Yesilyurt 12/13/2024

