

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
 Data File : VL041728.D
 Acq On : 11 Dec 2024 19:28
 Operator : SY/MD
 Sample : VL1211ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1211ABL01

Quant Time: Dec 12 13:39:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Dec 12 13:35:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.797	49	272196	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.972	114	1359828	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	1203529	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.390	95	788122	9.961	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
Data File : VL041728.D
Acq On : 11 Dec 2024 19:28
Operator : SY/MD
Sample : VL1211ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1211ABL01

Quant Time: Dec 12 13:39:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Dec 12 13:35:39 2024
Response via : Initial Calibration

