

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
 Data File : VL041737.D
 Acq On : 12 Dec 2024 00:13
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Dec 12 06:04:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Dec 12 05:50:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.780	49	252039	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	3.952	114	1228233	10.000	ppbv	-0.02
55) Chlorobenzene-d5	8.878	117	1082749	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.374	95	711354	9.994	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.900%
Target Compounds						
					Qvalue	
13) Ethanol	1.716	45	2269	1.875	ppbv #	34
15) Acetone	1.832	43	10990	0.378	ppbv #	86
19) Isopropyl Alcohol	1.884	45	15544	0.562	ppbv #	31
20) Methylene Chloride	2.059	84	5877	0.333	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121124\
Data File : VL041737.D
Acq On : 12 Dec 2024 00:13
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Dec 12 06:04:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121124AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Dec 12 05:50:25 2024
Response via : Initial Calibration

