

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
 Data File : VL039978.D
 Acq On : 12 Dec 2022 16:27
 Operator : SY/MD
 Sample : N5969-04DL 5X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DL

Quant Time: Dec 13 06:11:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.172	49	1136798	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.635	114	3099810	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.452	117	2822736	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	2280104	10.575	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.800%
Target Compounds						
					Qvalue	
9) Propene	2.687	41	15887	0.215	ppbv	84
13) Ethanol	3.237	45	20665	3.663	ppbv	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
Data File : VL039978.D
Acq On : 12 Dec 2022 16:27
Operator : SY/MD
Sample : N5969-04DL 5X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1DL

Quant Time: Dec 13 06:11:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

