

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
 Data File : VL039984.D
 Acq On : 12 Dec 2022 20:25
 Operator : SY/MD
 Sample : N5971-04DL 5X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-2DL

Quant Time: Dec 13 07:09:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.169	49	1138273	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.629	114	2965305	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.449	117	2712827	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.760	95	2179224	10.517	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.200%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.709	85	23437	0.144	ppbv #	76
4) Chloromethane	2.790	50	11001	0.110	ppbv	93
9) Propene	2.677	41	23046	0.311	ppbv	91
13) Ethanol	3.217	45	187934	33.269	ppbv	89
15) Acetone	3.452	43	157495	1.043	ppbv	95
20) Methylene Chloride	3.909	84	169065	2.311	ppbv	91
26) Hexane	5.192	57	91699	0.757	ppbv #	37

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
Data File : VL039984.D
Acq On : 12 Dec 2022 20:25
Operator : SY/MD
Sample : N5971-04DL 5X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-2DL

Quant Time: Dec 13 07:09:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

