

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
 Data File : VL039986.D
 Acq On : 12 Dec 2022 21:44
 Operator : SY/MD
 Sample : N5971-07DL 5X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/14/2022
 Supervised By :Mahesh Dadoda 12/14/2022

Quant Time: Dec 13 07:10:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.169	49	1101442	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.632	114	2916706	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.449	117	2626675	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.761	95	2130061	10.617	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.200%
Target Compounds						
					Qvalue	
13) Ethanol	3.214	45	494239	90.419	ppbv	99
15) Acetone	3.452	43	143811	0.984	ppbv	99
20) Methylene Chloride	3.912	84	340759	4.814	ppbv	94
26) Hexane	5.192	57	215587	1.840	ppbv #	37
53) Toluene	9.198	91	25247m	0.089	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
Data File : VL039986.D
Acq On : 12 Dec 2022 21:44
Operator : SY/MD
Sample : N5971-07DL 5X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-1DL

Quant Time: Dec 13 07:10:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/14/2022
Supervised By : Mahesh Dadoda 12/14/2022

