

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
 Data File : VL039990.D
 Acq On : 13 Dec 2022 00:23
 Operator : SY/MD
 Sample : N5971-09DL 5X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL

Quant Time: Dec 13 07:11:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.169	49	1084241	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.632	114	2823245	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.449	117	2496810	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.761	95	2078886	10.901	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	109.000%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.706	85	17403	0.112	ppbv	93
3) Chlorodifluoromethane	2.655	51	17005	0.114	ppbv	92
13) Ethanol	3.214	45	594875	110.556	ppbv	96
15) Acetone	3.452	43	171424	1.192	ppbv	97
20) Methylene Chloride	3.912	84	33479	0.481	ppbv	84
26) Hexane	5.195	57	28504	0.247	ppbv #	40
53) Toluene	9.201	91	33867	0.124	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
Data File : VL039990.D
Acq On : 13 Dec 2022 00:23
Operator : SY/MD
Sample : N5971-09DL 5X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL

Quant Time: Dec 13 07:11:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

