

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
 Data File : VL039992.D
 Acq On : 13 Dec 2022 1:39
 Operator : SY/MD
 Sample : N6020-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/14/2022
 Supervised By :Mahesh Dadoda 12/14/2022

Quant Time: Dec 13 07:12:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.172	49	1130463	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.632	114	3066307	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.449	117	2719110	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.757	95	2203909	10.612	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.100%
Target Compounds						
					Qvalue	
20) Methylene Chloride	3.915	84	25334	0.349	ppbv #	86
29) Chloroform	5.253	83	46917	0.294	ppbv	94
52) Tetrachloroethene	10.603	164	8358m	0.086	ppbv	
53) Toluene	9.195	91	19028m	0.064	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
Data File : VL039992.D
Acq On : 13 Dec 2022 1:39
Operator : SY/MD
Sample : N6020-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/14/2022
Supervised By :Mahesh Dadoda 12/14/2022

Quant Time: Dec 13 07:12:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

