

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
 Data File : VL039997.D
 Acq On : 13 Dec 2022 4:54
 Operator : SY/MD
 Sample : N5969-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-2DL

Quant Time: Dec 13 07:15:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.169	49	838756	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.632	114	2620038	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.449	117	2383534	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.757	95	1872223	10.284	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.800%
Target Compounds						
					Qvalue	
11) Trichlorofluoromethane	3.539	101	8988	0.061	ppbv	92
13) Ethanol	3.240	45	29125m	6.997	ppbv	
15) Acetone	3.452	43	862577	7.753	ppbv	100
36) Benzene	6.349	78	39458	0.177	ppbv #	93
52) Tetrachloroethene	10.597	164	91127	1.094	ppbv	90
62) Isopropylbenzene	14.011	105	74178	0.199	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
Data File : VL039997.D
Acq On : 13 Dec 2022 4:54
Operator : SY/MD
Sample : N5969-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-2DL

Quant Time: Dec 13 07:15:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

