

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
 Data File : VL039999.D
 Acq On : 13 Dec 2022 6:08
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Dec 13 07:15:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.172	49	886153	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.632	114	2556350	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.449	117	2324062	10.000	ppbv	0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.761	95	1783583	10.048	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121222\
Data File : VL039999.D
Acq On : 13 Dec 2022 6:08
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Dec 13 07:15:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

