

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121322\
 Data File : VL040010.D
 Acq On : 13 Dec 2022 17:36
 Operator : SY/MD
 Sample : N5969-01DL2 400X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SG-1DL2

Quant Time: Dec 14 02:38:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.172	49	1142282	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.632	114	3128838	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.452	117	2856571	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.760	95	2172606	9.957	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.600%
Target Compounds						
					Qvalue	
10) Heptane	7.561	43	278429	1.554	ppbv #	27
44) 2,2,4-Trimethylpentane	7.311	57	5661729	12.752	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121322\
Data File : VL040010.D
Acq On : 13 Dec 2022 17:36
Operator : SY/MD
Sample : N5969-01DL2 400X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-1DL2

Quant Time: Dec 14 02:38:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 02 23:21:36 2022
Response via : Initial Calibration

