

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL121421\
 Data File : VL038179.D
 Acq On : 15 Dec 2021 8:23
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Dec 15 13:35:15 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Nov 18 13:16:09 2021
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	811151	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.13	114	2612907	10.00	ppbv	-0.05
55) Chlorobenzene-d5	12.03	117	2266058	10.00	ppbv	-0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.36	95	1764857	10.80	ppbv	-0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.85	43	16000	0.228	ppbv #	78
20) Methylene Chloride	4.31	84	12329	0.313	ppbv #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL121421\
Data File : VL038179.D
Acq On : 15 Dec 2021 8:23
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Dec 15 13:35:15 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Nov 18 13:16:09 2021
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

