


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL121621\
 Data File : VL038183.D
 Acq On : 16 Dec 2021 10:31
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL111821AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Nov 18 13:16:09 2021
 Last Update : Thu Nov 18 13:16:09 2021



AutoFind: Scans 3611, 3612, 3613; Background Corrected with Scan 3579

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	23.8	172245	PASS
75	95	30	66	54.6	395370	PASS
95	95	100	100	100.0	723520	PASS
96	95	5	9	6.5	47245	PASS
173	174	0.00	2	1.0	3885	PASS
174	95	50	120	56.1	406250	PASS
175	174	4	9	7.7	31088	PASS
176	174	93	101	97.1	394538	PASS
177	176	5	9	6.3	24720	PASS