

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121624\
 Data File : VL041753.D
 Acq On : 16 Dec 2024 14:03
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Quant Time: Dec 19 06:12:42 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121624AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Thu Dec 19 06:09:59 2024

Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 12/20/2024
 Supervised By : Mahesh Dadoda 12/20/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.787	49	67443	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	182992	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.882	117	163971	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.377 95 138312 10.122 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 101.200%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.499	85	24819	2.068	ppbv	98
3) Chlorodifluoromethane	1.476	51	22299	2.027	ppbv	96
4) Chloromethane	1.531	50	9926	2.114	ppbv	99
5) Vinyl Chloride	1.580	62	9300	1.966	ppbv	99
6) Bromomethane	1.667	94	4625	2.072	ppbv	99
7) Chloroethane	1.706	64	4003	2.178	ppbv #	85
8) Dichlorotetrafluoroethane	1.554	85	21791	2.029	ppbv	100
9) Propene	1.483	41	10115	2.122	ppbv	97
10) Heptane	4.978	43	25297	2.066	ppbv	100
11) Trichlorofluoromethane	1.878	101	26208	2.056	ppbv	96
12) 1,1,2-Trichlorotrifluo...	2.140	101	19770	2.039	ppbv	99
13) Ethanol	1.722	45	792	1.998	ppbv #	49
14) Bromoethene	1.784	108	6527	1.994	ppbv #	90
15) Acetone	1.835	43	27613	2.104	ppbv	99
16) 1,3-Butadiene	1.609	39	10478	2.028	ppbv	96
17) tert-Butyl alcohol	2.043	59	26646	2.047	ppbv	99
18) 1,1-Dichloroethene	2.033	96	8389	2.064	ppbv	92
19) Isopropyl Alcohol	1.887	45	14937	1.897	ppbv	93
20) Methylene Chloride	2.062	84	7586	2.027	ppbv	94
21) Allyl Chloride	2.094	41	14731	2.070	ppbv	99
22) trans-1,2-Dichloroethene	2.340	96	9407	2.022	ppbv	94
23) Vinyl Acetate	2.463	43	17652	3.096	ppbv #	95
24) 1,1-Dichloroethane	2.408	63	19070	2.025	ppbv	99
25) Ethyl Acetate	2.836	43	50510	2.059	ppbv	99
26) Hexane	2.826	57	19200	2.073	ppbv	98
27) Carbon Disulfide	2.153	76	16214	1.942	ppbv #	34
28) Methyl tert-Butyl Ether	2.441	73	15039	2.001	ppbv	98
29) Chloroform	2.849	83	28843	2.043	ppbv	98
30) Cyclohexane	3.858	84	16384	2.079	ppbv	99
31) cis-1,2-Dichloroethene	2.719	61	19688	2.057	ppbv	96
32) 1,1,1-Trichloroethane	3.370	97	28341	2.017	ppbv	98
34) 2-Butanone	2.557	43	31483	2.138	ppbv	99
35) Carbon Tetrachloride	3.765	117	27727	2.134	ppbv	98
36) Benzene	3.658	78	38759	2.048	ppbv	94
37) 1,2-Dichloroethane	3.221	62	23800	2.021	ppbv	98
38) Trichloroethene	4.548	130	15660	1.965	ppbv	94
39) 1,2-Dichloropropane	4.315	63	14197	2.074	ppbv	94
40) 1,4-Dioxane	4.603	88	6826m	2.041	ppbv	
41) Tetrahydrofuran	3.059	42	16665	2.021	ppbv	96
42) Bromodichloromethane	4.489	83	26716	1.972	ppbv	98
43) Methyl Methacrylate	4.884	69	14849	2.059	ppbv	100
44) 2,2,4-Trimethylpentane	4.642	57	67432	2.086	ppbv	100
45) t-1,3-Dichloropropene	6.415	75	14885	1.995	ppbv	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121624\
 Data File : VL041753.D
 Acq On : 16 Dec 2024 14:03
 Operator : SY/MD
 Sample : VSTDICC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC002

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/20/2024
 Supervised By :Mahesh Dadoda 12/20/2024

Quant Time: Dec 19 06:12:42 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121624AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Thu Dec 19 06:09:59 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.606	75	19105m	2.017	ppbv	
47) 1,1,2-Trichloroethane	6.584	97	14858	2.078	ppbv	89
48) Dibromochloromethane	7.390	129	19764	2.024	ppbv	98
49) Bromoform	9.484	173	16200	2.027	ppbv	98
50) 4-Methyl-2-Pentanone	5.758	43	42045	2.098	ppbv	97
51) 2-Hexanone	7.441	43	34878	2.147	ppbv #	98
52) Tetrachloroethene	8.225	164	13612	1.932	ppbv	93
53) Toluene	6.933	91	44470	2.072	ppbv	99
54) 1,2-Dibromoethane	7.655	107	19820	2.012	ppbv	91
56) 1,1,1,2-Tetrachloroethane	8.924	131	18613	2.141	ppbv	98
57) Chlorobenzene	8.920	112	35685	2.127	ppbv	99
58) Ethyl Benzene	9.354	91	63316	2.145	ppbv	97
59) m/p-Xylene	9.535	91	99821m	4.204	ppbv	
60) o-Xylene	9.966	91	51152	2.149	ppbv	99
61) Styrene	9.872	104	20842	2.025	ppbv	96
62) Isopropylbenzene	10.539	105	75206	2.103	ppbv	100
63) 1,1,2,2-Tetrachloroethane	9.966	83	29751	1.997	ppbv	100
64) n-propylbenzene	10.992	120	19146	2.117	ppbv	99
65) tert-Butylbenzene	11.532	119	68879	2.135	ppbv	98
66) Benzyl Chloride	11.617	91	5400	1.898	ppbv	99
67) sec-Butylbenzene	11.769	105	95968	2.149	ppbv	100
69) p-Isopropyltoluene	11.927	119	79236	2.153	ppbv	100
70) n-Butylbenzene	12.283	91	75430	2.122	ppbv	99
71) 2-Chlorotoluene	10.901	91	57724	2.114	ppbv	99
72) 4-Ethyltoluene	11.125	105	58841	2.098	ppbv	99
73) 1,3,5-Trimethylbenzene	11.206	105	48645	2.087	ppbv	97
74) 1,2,4-Trimethylbenzene	11.539	105	53084	2.118	ppbv	96
75) 1,3-Dichlorobenzene	11.610	146	32609	2.040	ppbv	98
76) 1,4-Dichlorobenzene	11.675	146	32527	2.060	ppbv	98
77) 1,2-Dichlorobenzene	11.950	146	32223	2.095	ppbv	100
78) Hexachloro-1,3-Butadiene	13.882	225	25358	2.224	ppbv	98
79) Naphthalene	13.513	128	37855	2.167	ppbv	100
80) Naphthalene,2-methyl-	14.462	142	4482	0.830	ppbv	94
81) 1,2,4-Trichlorobenzene	13.442	180	19425	2.035	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

