

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL121718\
 Data File : VL033005.D
 Acq On : 17 Dec 2018 21:12
 Operator : SY/AP
 Sample : J6382-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 P0131-IA001-181212-01DL

Manual Integrations
 APPROVED

MMDadoda
 12/19/2018 2:00:18 PM

Quant Time: Dec 18 02:22:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec 0
 QLast Update : Fri Dec 07 09:35:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1313975	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3181562	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3464299	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.59	95	2504277	9.78	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
4) Chloromethane	3.18	50	11689m	0.202	ppbv	
13) Ethanol	3.67	45	16497	1.218	ppbv	95
15) Acetone	3.93	43	44768	0.397	ppbv	95
20) Methylene Chloride	4.42	84	24265	0.440	ppbv #	94
26) Hexane	5.79	57	18678	0.152	ppbv #	54
53) Toluene	9.95	91	120841	0.487	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL121718\
Data File : VL033005.D
Acq On : 17 Dec 2018 21:12
Operator : SY/AP
Sample : J6382-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
P0131-IA001-181212-01DL

Manual Integrations
APPROVED

MMDadoda
12/19/2018 2:00:18 PM

Quant Time: Dec 18 02:22:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL120718AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Dec
QLast Update : Fri Dec 07 09:35:17 2018
Response via : Initial Calibration

