


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL121719\
 Data File : VL034492.D
 Acq On : 17 Dec 2019 12:37
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL111919AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Nov 19 14:15:35 2019
 Last Update : Tue Nov 19 14:15:35 2019



AutoFind: Scans 3633, 3634, 3635; Background Corrected with Scan 3609

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	26.5	179264	PASS
75	95	30	66	63.6	429525	PASS
95	95	100	100	100.0	675669	PASS
96	95	5	9	6.5	43915	PASS
173	174	0.00	2	0.6	2578	PASS
174	95	50	120	62.7	423808	PASS
175	174	4	9	7.8	33080	PASS
176	174	93	101	96.6	409429	PASS
177	176	5	9	7.2	29362	PASS