

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121724\
Data File : VL041781.D
Acq On : 17 Dec 2024 16:17
Operator : SY/MD
Sample : MDL 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL 0.03

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/18/2024
Supervised By :Semsettin Yesilyurt 12/18/2024

Quant Time: Dec 21 03:06:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121624AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
QLast Update : Thu Dec 19 07:32:09 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	51238	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	131572	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	111343	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.383	95	93730	10.102	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.000%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	145	0.040	ppbv #	87
32) 1,1,1-Trichloroethane	3.373	97	359	0.033	ppbv #	67
35) Carbon Tetrachloride	3.780	117	337m	0.036	ppbv	
38) Trichloroethene	4.564	130	82m	0.014	ppbv	
52) Tetrachloroethene	8.221	164	156m	0.031	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.962	83	390m	0.039	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121724\
Data File : VL041781.D
Acq On : 17 Dec 2024 16:17
Operator : SY/MD
Sample : MDL 0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MDL 0.03

Quant Time: Dec 21 03:06:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121624AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Dec 19 07:32:09 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 12/18/2024
Supervised By : Semsettin Yesilyurt 12/18/2024

