

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121922\
 Data File : VL040024.D
 Acq On : 19 Dec 2022 12:06
 Operator : SY/MD
 Sample : VSTDIC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/20/2022
 Supervised By :Mahesh Dadoda 12/20/2022

Quant Time: Dec 20 04:41:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Tue Dec 20 04:20:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.173	49	998638	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.636	114	2871898	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.455	117	2482394	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.761	95	1941869	10.241	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.400%
Target Compounds						
5) Vinyl Chloride	2.903	62	7411	0.099	ppbv	90
32) 1,1,1-Trichloroethane	5.989	97	16420m	0.089	ppbv	
35) Carbon Tetrachloride	6.475	117	15150m	0.085	ppbv	
38) Trichloroethene	7.269	130	11659m	0.088	ppbv	
52) Tetrachloroethene	10.597	164	8753m	0.100	ppbv	
54) 1,2-Dibromoethane	9.996	107	8802m	0.078	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.040	83	10909m	0.101	ppbv	
79) Naphthalene	21.387	128	11368m	0.059	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121922\
Data File : VL040024.D
Acq On : 19 Dec 2022 12:06
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/20/2022
Supervised By :Mahesh Dadoda 12/20/2022

Quant Time: Dec 20 04:41:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Dec 20 04:20:10 2022
Response via : Initial Calibration

