

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121922\
 Data File : VL040025.D
 Acq On : 19 Dec 2022 12:43
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/20/2022
 Supervised By :Mahesh Dadoda 12/20/2022

Quant Time: Dec 20 04:23:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Tue Dec 20 04:20:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	964803	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.635	114	2821156	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.452	117	2410030	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	1867975	10.147	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.500%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.912	62	1781m	0.025	ppbv	
32) 1,1,1-Trichloroethane	5.986	97	4428m	0.025	ppbv	
35) Carbon Tetrachloride	6.462	117	4254m	0.024	ppbv	
38) Trichloroethene	7.272	130	2689m	0.021	ppbv	
52) Tetrachloroethene	10.609	164	2626m	0.030	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.040	83	2719m	0.026	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121922\
Data File : VL040025.D
Acq On : 19 Dec 2022 12:43
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Dec 20 04:23:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Dec 20 04:20:10 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/20/2022
Supervised By : Mahesh Dadoda 12/20/2022

