

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121922\
 Data File : VL040035.D
 Acq On : 19 Dec 2022 19:31
 Operator : SY/MD
 Sample : N6059-03DL 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/20/2022
 Supervised By :Mahesh Dadoda 12/20/2022

Quant Time: Dec 20 05:15:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Tue Dec 20 04:50:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.169	49	994821	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.629	114	2845380	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.446	117	2415819	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.757	95	1858304	10.070	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.700%
Target Compounds						
					Qvalue	
9) Propene	2.684	41	10414	0.167	ppbv #	82
13) Ethanol	3.231	45	36455m	6.218	ppbv	
15) Acetone	3.462	43	69000	0.448	ppbv	94
19) Isopropyl Alcohol	3.571	45	79028	1.042	ppbv #	93
20) Methylene Chloride	3.912	84	100139	1.531	ppbv	95
26) Hexane	5.192	57	41938	0.374	ppbv #	40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL121922\
Data File : VL040035.D
Acq On : 19 Dec 2022 19:31
Operator : SY/MD
Sample : N6059-03DL 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/20/2022
Supervised By :Mahesh Dadoda 12/20/2022

Quant Time: Dec 20 05:15:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Dec 20 04:50:21 2022
Response via : Initial Calibration

