

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122023\
 Data File : VL040919.D
 Acq On : 20 Dec 2023 14:17
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 12/21/2023
 Supervised By : Mahesh Dadoda 12/21/2023

Quant Time: Dec 21 07:11:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Dec 21 07:06:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.182	49	1249503	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.635	114	3035543	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.442	117	2603299	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.751	95	1934033	10.086	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.900%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.928	62	1844m	0.029	ppbv	
32) 1,1,1-Trichloroethane	5.996	97	5502m	0.030	ppbv	
35) Carbon Tetrachloride	6.484	117	5419m	0.028	ppbv	
38) Trichloroethene	7.272	130	2804m	0.031	ppbv	
52) Tetrachloroethene	10.584	164	2169m	0.027	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.031	83	5832m	0.030	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122023\
Data File : VL040919.D
Acq On : 20 Dec 2023 14:17
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Dec 21 07:11:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL122023AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Dec 21 07:06:17 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/21/2023
Supervised By :Mahesh Dadoda 12/21/2023

