

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL122121\
 Data File : VL038224.D
 Acq On : 21 Dec 2021 22:16
 Operator : SY/AP
 Sample : M5118-01DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-2DL

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/22/2021
 Supervised By :Mahesh Dadoda 12/29/2021

Quant Time: Dec 22 08:41:45 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL122121A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Dec 21 17:47:08 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	741404	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	7.19	114	2134452	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	1816861	10.00	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	1446541	10.33	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.04	41	10335m	0.191	ppbv	
13) Ethanol	3.64	45	34406m	3.752	ppbv	
15) Acetone	3.88	43	45401m	0.610	ppbv	
19) Isopropyl Alcohol	4.01	45	25158m	0.542	ppbv	
20) Methylene Chloride	4.36	84	83132	1.979	ppbv	98
26) Hexane	5.70	57	44063	0.436	ppbv #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL122121\
Data File : VL038224.D
Acq On : 21 Dec 2021 22:16
Operator : SY/AP
Sample : M5118-01DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-2DL

Manual Integrations
APPROVED

Quant Time: Dec 22 08:41:45 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL122121\
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Dec 21 17:47:08 2021
Response via : Initial Calibration

Reviewed By :Amit Patel 12/22/2021
Supervised By :Mahesh Dadoda 12/29/2021

