

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL122121\
 Data File : VL038228.D
 Acq On : 22 Dec 2021 00:54
 Operator : SY/AP
 Sample : M5118-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-1DL

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/22/2021
 Supervised By :Mahesh Dadoda 12/29/2021

Quant Time: Dec 22 08:50:51 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL122121A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Dec 21 17:47:08 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	747379	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.19	114	2023208	10.00	ppbv	0.02
55) Chlorobenzene-d5	12.10	117	1759117	10.00	ppbv	0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	1407729	10.38	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%

Target Compounds

					Ovalue
9) Propene	3.04	41	8237	0.151 ppbv	92
13) Ethanol	3.65	45	13916m	1.505 ppbv	
15) Acetone	3.89	43	40405m	0.539 ppbv	
19) Isopropyl Alcohol	4.03	45	4824m	0.103 ppbv	
20) Methylene Chloride	4.37	84	125530	2.964 ppbv	95
26) Hexane	5.71	57	45392	0.445 ppbv #	43

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL122121\
Data File : VL038228.D
Acq On : 22 Dec 2021 00:54
Operator : SY/AP
Sample : M5118-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-1DL

Quant Time: Dec 22 08:50:51 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL122121
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Dec 21 17:47:08 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 12/22/2021
Supervised By :Mahesh Dadoda 12/29/2021

