

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL122320\
 Data File : VL036197.D
 Acq On : 23 Dec 2020 18:11
 Operator : SY/AP
 Sample : L5196-01DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SSSG2DL

Manual Integrations
 APPROVED

MMDadoda
 12/26/2020 1:36:15 PM

Quant Time: Dec 24 05:49:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL120720AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 0
 QLast Update : Mon Dec 07 16:56:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	1212878	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.15	114	2467844	10.00	ppbv	-0.05
55) Chlorobenzene-d5	12.06	117	2218265	10.00	ppbv	-0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	1861061	10.57	ppbv	-0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	11020m	0.159	ppbv	
13) Ethanol	3.60	45	24878m	2.603	ppbv	
15) Acetone	3.84	43	62950m	0.432	ppbv	
20) Methylene Chloride	4.32	84	54536	0.913	ppbv #	80
26) Hexane	5.66	57	115070	0.878	ppbv #	40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL122320\
Data File : VL036197.D
Acq On : 23 Dec 2020 18:11
Operator : SY/AP
Sample : L5196-01DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SSSG2DL

Manual Integrations
APPROVED

MMDadoda
12/26/2020 1:36:15 PM

Quant Time: Dec 24 05:49:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL120720AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec
QLast Update : Mon Dec 07 16:56:49 2020
Response via : Initial Calibration

