


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL122320\
 Data File : VL036191.D
 Acq On : 23 Dec 2020 9:58
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL120720AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Dec 07 16:56:49 2020
 Last Update : Mon Dec 07 16:56:49 2020



AutoFind: Scans 3607, 3608, 3609; Background Corrected with Scan 3582

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	28.4	263061	PASS
75	95	30	66	62.3	578325	PASS
95	95	100	100	100.0	927808	PASS
96	95	5	9	6.7	61762	PASS
173	174	0.00	2	1.2	6020	PASS
174	95	50	120	55.3	513438	PASS
175	174	4	9	7.2	36933	PASS
176	174	93	101	97.6	501141	PASS
177	176	5	9	6.0	30069	PASS