

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122921\
 Data File : VL038248.D
 Acq On : 30 Dec 2021 1:11
 Operator : SY/AP
 Sample : M5252-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-2DL

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/30/2021
 Supervised By :Semsettin Yesilyurt 01/02/2022

Quant Time: Dec 30 06:20:56 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL122121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Dec 21 17:47:08 2021
 QLast Update : Tue Dec 21 17:47:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.671	49	685461	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.176	114	1692027	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.086	117	1428015	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.417	95	1134562	10.310	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.100%
Target Compounds						
9) Propene	3.034	41	14940	0.299	ppbv	87
13) Ethanol	3.604	45	1083415	127.773	ppbv #	100
15) Acetone	3.858	43	629974	9.157	ppbv	95
19) Isopropyl Alcohol	3.973	45	450029	10.488	ppbv #	39
20) Methylene Chloride	4.353	84	52983	1.364	ppbv #	84
26) Hexane	5.690	57	41614	0.445	ppbv #	53
38) Trichloroethene	7.838	130	29905	0.498	ppbv	95
52) Tetrachloroethene	11.224	164	15044m	0.258	ppbv	
59) m/p-Xylene	12.963	91	39103m	0.206	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122921\
Data File : VL038248.D
Acq On : 30 Dec 2021 1:11
Operator : SY/AP
Sample : M5252-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-2DL

Manual Integrations
APPROVED

Reviewed By :Amit Patel 12/30/2021
Supervised By :Semsettin Yesilyurt 01/02/2022

Quant Time: Dec 30 06:20:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL122121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Dec
QLast Update : Tue Dec 21 17:47:08 2021
Response via : Initial Calibration

