

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122921\
 Data File : VL038250.D
 Acq On : 30 Dec 2021 2:30
 Operator : SY/AP
 Sample : M5252-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DL

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/30/2021
 Supervised By :Semsettin Yesilyurt 01/02/2022

Quant Time: Dec 30 06:21:16 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL122121AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LTue Dec 2
 QLast Update : Tue Dec 21 17:47:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.674	49	654757	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.179	114	1608054	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.085	117	1376524	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.416	95	1063149	10.023	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.200%
Target Compounds						
9) Propene	3.031	41	12772m	0.267	ppbv	Qvalue
13) Ethanol	3.636	45	17454m	2.155	ppbv	
15) Acetone	3.874	43	38710m	0.589	ppbv	
20) Methylene Chloride	4.356	84	37118	1.000	ppbv	95
26) Hexane	5.693	57	30903	0.346	ppbv #	42

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL122921\
Data File : VL038250.D
Acq On : 30 Dec 2021 2:30
Operator : SY/AP
Sample : M5252-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1DL

Quant Time: Dec 30 06:21:16 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL122121AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Dec
QLast Update : Tue Dec 21 17:47:08 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 12/30/2021
Supervised By :Semsettin Yesilyurt 01/02/2022

