

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL123019\
 Data File : VL034514.D
 Acq On : 30 Dec 2019 13:35
 Operator : SY/AP
 Sample : VSTDIC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.03

Manual Integrations
 APPROVED

apatel
 1/2/2020 10:48:53 AM

Quant Time: Dec 30 14:43:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 30 2019
 QLast Update : Mon Dec 30 14:19:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1044838	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	2060237	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	2026707	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1698465	9.96	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.23	62	2955	0.041	ppbv	# 79
32) 1,1,1-Trichloroethane	6.52	97	6245m	0.036	ppbv	
35) Carbon Tetrachloride	7.03	117	6707m	0.038	ppbv	
38) Trichloroethene	7.87	130	2164m	0.026	ppbv	
52) Tetrachloroethene	11.27	164	2487m	0.028	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.73	83	7523m	0.037	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL123019\
Data File : VL034514.D
Acq On : 30 Dec 2019 13:35
Operator : SY/AP
Sample : VSTDIC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.03

Manual Integrations
APPROVED

apatel
1/2/2020 10:48:53 AM

Quant Time: Dec 30 14:43:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec
QLast Update : Mon Dec 30 14:19:57 2019
Response via : Initial Calibration

