

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL123019\
 Data File : VL034521.D
 Acq On : 30 Dec 2019 21:09
 Operator : SY/AP
 Sample : K6388-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EFFLUENTDL

Manual Integrations
 APPROVED

apatel
 1/2/2020 10:49:28 AM

Quant Time: Dec 31 05:41:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 31 05:41:46 2019
 QLast Update : Mon Dec 30 14:58:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	993293	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	1957971	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	1945948	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1631026	10.04	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
7) Chloroethane	3.53	64	7670m	0.320	ppbv	
9) Propene	3.00	41	8595m	0.188	ppbv	
13) Ethanol	3.60	45	4229m	0.334	ppbv	
15) Acetone	3.85	43	36678	0.253	ppbv	92
18) 1,1-Dichloroethene	4.28	96	6583m	0.131	ppbv	
19) Isopropyl Alcohol	3.98	45	10614m	0.113	ppbv	
20) Methylene Chloride	4.33	84	15450	0.271	ppbv	87
24) 1,1-Dichloroethane	5.00	63	23065m	0.172	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL123019\
Data File : VL034521.D
Acq On : 30 Dec 2019 21:09
Operator : SY/AP
Sample : K6388-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EFFLUENTDL

Manual Integrations
APPROVED

apatel
1/2/2020 10:49:28 AM

Quant Time: Dec 31 05:41:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec
QLast Update : Mon Dec 30 14:58:50 2019
Response via : Initial Calibration

