

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL123019\
 Data File : VL034523.D
 Acq On : 30 Dec 2019 22:28
 Operator : SY/AP
 Sample : K6388-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 MIDDLE

Manual Integrations
 APPROVED

apatel
 1/2/2020 10:49:35 AM

Quant Time: Dec 31 05:49:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Dec 31 05:49:50 2019
 QLast Update : Mon Dec 30 14:58:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	955199	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	1919943	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	1899015	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1594016	10.06	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	6258m	0.143	ppbv	
13) Ethanol	3.60	45	3100m	0.255	ppbv	
15) Acetone	3.84	43	39572	0.284	ppbv	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL123019\
Data File : VL034523.D
Acq On : 30 Dec 2019 22:28
Operator : SY/AP
Sample : K6388-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
MIDDLE

Manual Integrations
APPROVED

apatel
1/2/2020 10:49:35 AM

Quant Time: Dec 31 05:49:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec
QLast Update : Mon Dec 30 14:58:50 2019
Response via : Initial Calibration

