

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL123019\
 Data File : VL034528.D
 Acq On : 31 Dec 2019 1:46
 Operator : SY/AP
 Sample : OC121819 CAN 10269
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC121819 CAN 10269

Quant Time: Dec 31 14:35:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 30 14:58:50 2019
 QLast Update : Mon Dec 30 14:58:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	928122	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	1813430	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	1813290	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1481926	9.79	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL123019\
Data File : VL034528.D
Acq On : 31 Dec 2019 1:46
Operator : SY/AP
Sample : OC121819 CAN 10269
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC121819 CAN 10269

Quant Time: Dec 31 14:35:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Dec 30 14:58:50 2019
QLast Update : Mon Dec 30 14:58:50 2019
Response via : Initial Calibration

