


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL123019\
 Data File : VL034508.D
 Acq On : 30 Dec 2019 9:45
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL123019AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Nov 19 14:15:35 2019
 Last Update : Tue Nov 19 14:15:35 2019



AutoFind: Scans 3629, 3630, 3631; Background Corrected with Scan 3606

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	27.8	211882	PASS
75	95	30	66	63.8	485449	PASS
95	95	100	100	100.0	761173	PASS
96	95	5	9	6.7	51349	PASS
173	174	0.00	2	0.8	4112	PASS
174	95	50	120	66.0	502613	PASS
175	174	4	9	6.9	34682	PASS
176	174	93	101	94.8	476650	PASS
177	176	5	9	6.7	32053	PASS