

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030322\
 Data File : VL038524.D
 Acq On : 3 Mar 2022 21:50
 Operator : SY/AP
 Sample : CAN 10326
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10326

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/05/2022
 Supervised By :Mahesh Dadoda 03/07/2022

Quant Time: Mar 04 06:29:34 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022222A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Feb 22 19:45:51 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	780478	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.14	114	1687862	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.05	117	1236868	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	1048977	10.25	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	8687m	1.913	ppbv	
15) Acetone	3.84	43	26930m	0.406	ppbv	
20) Methylene Chloride	4.33	84	89746	2.729	ppbv	95
26) Hexane	5.66	57	50786	0.629	ppbv #	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030322\
Data File : VL038524.D
Acq On : 3 Mar 2022 21:50
Operator : SY/AP
Sample : CAN 10326
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10326

Manual Integrations
APPROVED

Quant Time: Mar 04 06:29:34 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022222
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 03/05/2022
Supervised By :Mahesh Dadoda 03/07/2022

