

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030322\
Data File : VL038528.D
Acq On : 4 Mar 2022 00:27
Operator : SY/AP
Sample : CAN 10595
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10595

Quant Time: Mar 04 06:44:11 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	807320	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.14	114	1812060	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.05	117	1597042	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.37	95	1324868	10.03	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.30%


Target Compounds	Ovalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030322\
Data File : VL038528.D
Acq On : 4 Mar 2022 00:27
Operator : SY/AP
Sample : CAN 10595
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10595

Quant Time: Mar 04 06:44:11 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.64 Instrument:

Delta R.T. MSVOA_1

Lab File: ClientSampleId:

Acq: 4 Mar 2022 00:27


Tot Ion: 49 Resp: 807320

Ion Ratio Lower Upper

49 100

128 43.6 24.3 72.9




130 58.6 30.1 90.5



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): V

Ion 130.00 (129.70 to 130.70): V



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.14 min Scan# 1353

Delta R.T. -0.02 min

Lab File: VL038528.D

Acq: 4 Mar 2022 00:27


Tgt Ion:114 Resp: 1812060

Ion Ratio Lower Upper

114 100

63 29.1 22.2 33.2



88 19.3 15.2 22.8




Abundance Ion 114.10 (113.80 to 114.80): V

Ion 63.10 (62.80 to 63.80): VL0





Ion 88.10 (87.80 to 88.80): VL0





#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.05 min
Delta R.T. MSVOA_L
Lab File: ClientSampleId :
Acq: 4 Mar 2022 00:27

Tot Ion:117 Resp: 1597042
Ion Ratio Lower Upper
117 100
82 74.3 54.9 82.3
119 32.7 46.1 69.1#



#68
1-Bromo-4-Fluorobenzene
Concen: 10.026 ppbv
RT: 14.37 min Scan# 3603
Delta R.T. -0.02 min
Lab File: VL038528.D
Acq: 4 Mar 2022 00:27

Tgt Ion: 95 Resp: 1324868
Ion Ratio Lower Upper
95 100
174 58.2 49.0 73.4
175 4.3 3.6 5.4

