

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010325\
 Data File : VL041829.D
 Acq On : 03 Jan 2025 15:48
 Operator : SY/MD
 Sample : VL0103ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0103ABL01

Quant Time: Jan 03 16:11:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121624AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Dec 19 07:32:09 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.803	49	85723	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	3.981	114	203573	10.000	ppbv	0.03
55) Chlorobenzene-d5	8.904	117	176522	10.000	ppbv	0.03

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.396	95	142384	9.679	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010325\
Data File : VL041829.D
Acq On : 03 Jan 2025 15:48
Operator : SY/MD
Sample : VL0103ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0103ABL01

Quant Time: Jan 03 16:11:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121624AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Dec 19 07:32:09 2024
Response via : Initial Calibration

