

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
 Data File : VL040083.D
 Acq On : 9 Jan 2023 19:43
 Operator : SY/MD
 Sample : QC010523 CAN 10279
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC010523 CAN 10279

Quant Time: Jan 10 01:22:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jan 10 00:22:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.188	49	868848	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.651	114	2363655	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.471	117	2046702	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.783	95	1644781	10.520	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
Data File : VL040083.D
Acq On : 9 Jan 2023 19:43
Operator : SY/MD
Sample : QC010523 CAN 10279
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC010523 CAN 10279

Quant Time: Jan 10 01:22:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Jan 10 00:22:08 2023
Response via : Initial Calibration

