

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
 Data File : VL040085.D
 Acq On : 9 Jan 2023 23:14
 Operator : SY/MD
 Sample : QC122822 CAN 10269
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC122822 CAN 10269

Quant Time: Jan 11 01:09:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jan 10 00:22:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	919631	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	2312160	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.468	117	1994022	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.783	95	1612813	10.588	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.900%
Target Compounds						
						Qvalue
20) Methylene Chloride	3.925	84	52942	0.876	ppbv	97
26) Hexane	5.205	57	38669	0.373	ppbv #	35

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
Data File : VL040085.D
Acq On : 9 Jan 2023 23:14
Operator : SY/MD
Sample : QC122822 CAN 10269
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC122822 CAN 10269

Quant Time: Jan 11 01:09:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Jan 10 00:22:08 2023
Response via : Initial Calibration

