

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
 Data File : VL040088.D
 Acq On : 10 Jan 2023 8:59
 Operator : SY/MD
 Sample : QC122822 CAN 10269
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC122822 CAN 10269

Quant Time: Jan 11 01:10:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jan 10 00:22:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.179	49	916061	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	2400775	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2114689	10.000	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.777	95	1750340	10.836	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
Data File : VL040088.D
Acq On : 10 Jan 2023 8:59
Operator : SY/MD
Sample : QC122822 CAN 10269
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC122822 CAN 10269

Quant Time: Jan 11 01:10:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Jan 10 00:22:08 2023
Response via : Initial Calibration

