

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
 Data File : VL040089.D
 Acq On : 10 Jan 2023 9:36
 Operator : SY/MD
 Sample : QC122922 CAN 10609
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC122922 CAN 10609

Quant Time: Jan 11 01:10:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Jan 10 00:22:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.182	49	903717	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.642	114	2361676	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	2098149	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.773	95	1733603	10.817	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.200%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010923\
Data File : VL040089.D
Acq On : 10 Jan 2023 9:36
Operator : SY/MD
Sample : QC122922 CAN 10609
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC122922 CAN 10609

Quant Time: Jan 11 01:10:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL121922AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Jan 10 00:22:08 2023
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 10.000 ppbv
RT: 5.182 min Scan# 802
Delta R.T. -0.003 min
Lab File: VL040089.D
Acq: 10 Jan 2023 9:36




Instrument :
MSVOA_L
ClientSampleId :
QC122922 CAN 10609


Tgt Ion: 49 Resp: 903717
Ion Ratio Lower Upper
49 100
128 47.1 22.4 67.3
130 60.1 28.9 86.8



#33
1,4-Difluorobenzene
Concen: 10.000 ppbv
RT: 6.642 min Scan# 1256
Delta R.T. -0.010 min
Lab File: VL040089.D
Acq: 10 Jan 2023 9:36

Tgt Ion: 114 Resp: 2361676
Ion Ratio Lower Upper
114 100
63 26.6 20.2 30.2
88 18.4 14.6 21.8









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 11.465 min Scan# 21
Delta R.T. -0.010 min
Lab File: VL040089.D
Acq: 10 Jan 2023 9:36

Instrument :
MSVOA_L
ClientSampleId :
QC122922 CAN 10609

Tgt Ion:117 Resp: 2098149
Ion Ratio Lower Upper
117 100
82 68.9 51.2 76.8
119 32.8 23.2 34.8



#68
1-Bromo-4-Fluorobenzene
Concen: 10.817 ppbv
RT: 13.773 min Scan# 3474
Delta R.T. -0.013 min
Lab File: VL040089.D
Acq: 10 Jan 2023 9:36

Tgt Ion: 95 Resp: 1733603
Ion Ratio Lower Upper
95 100
174 59.6 48.6 73.0
175 4.2 3.4 5.2

