

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010925\
 Data File : VL041864.D
 Acq On : 09 Jan 2025 13:54
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 01/10/2025
 Supervised By :Semsettin Yesilyurt 01/10/2025

Quant Time: Jan 09 14:27:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL010925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Jan 09 13:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	140652	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.982	114	417764	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.901	117	382336	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.393	95	308087	9.408	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.100%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.590	62	375	0.039	ppbv #	86
32) 1,1,1-Trichloroethane	3.389	97	1032m	0.041	ppbv	
35) Carbon Tetrachloride	3.791	117	1177m	0.037	ppbv	
38) Trichloroethene	4.580	130	518m	0.033	ppbv	
52) Tetrachloroethene	8.251	164	686	0.058	ppbv #	49
63) 1,1,2,2-Tetrachloroethane	9.979	83	1052	0.027	ppbv #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010925\
Data File : VL041864.D
Acq On : 09 Jan 2025 13:54
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/10/2025
Supervised By :Semsettin Yesilyurt 01/10/2025

Quant Time: Jan 09 14:27:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL010925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Thu Jan 09 13:37:17 2025
Response via : Initial Calibration

