

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL011321\
 Data File : VL036215.D
 Acq On : 13 Jan 2021 5:43
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

MMDadoda
 1/18/2021 3:48:57 PM

Quant Time: Jan 14 16:13:38 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Jan 13
 QLast Update : Wed Jan 13 05:46:16 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1550419	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3372221	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3026774	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2462001	10.14	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.26	62	3133m	0.033	ppbv	
32) 1,1,1-Trichloroethane	6.50	97	7184m	0.032	ppbv	
35) Carbon Tetrachloride	7.02	117	6729m	0.034	ppbv	
38) Trichloroethene	7.83	130	4869m	0.046	ppbv	
52) Tetrachloroethene	11.21	164	5244m	0.051	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.70	83	10258m	0.046	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL011321\
Data File : VL036215.D
Acq On : 13 Jan 2021 5:43
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
1/18/2021 3:48:57 PM

Quant Time: Jan 14 16:13:38 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Jan
QLast Update : Wed Jan 13 05:46:16 2021
Response via : Initial Calibration

