

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL011321\
 Data File : VL036221.D
 Acq On : 13 Jan 2021 11:57
 Operator : SY/AP
 Sample : M1066-02DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 0121-SS2DL

Manual Integrations
 APPROVED

MMDadoda
 1/18/2021 3:49:16 PM

Quant Time: Jan 16 22:01:15 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LThu Jan 14 16:13:38 2021
 QLast Update : Thu Jan 14 16:13:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1538679	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	3287964	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2913196	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	2483578	10.30	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.02	41	125102	1.390	ppbv	# 80
13) Ethanol	3.61	45	1200579m	84.965	ppbv	
15) Acetone	3.85	43	877043	4.654	ppbv	96
17) tert-Butyl alcohol	4.30	59	282915m	1.819	ppbv	
53) Toluene	9.81	91	68204m	0.202	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL011321\
Data File : VL036221.D
Acq On : 13 Jan 2021 11:57
Operator : SY/AP
Sample : M1066-02DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
0121-SS2DL

Manual Integrations
APPROVED

MMDadoda
1/18/2021 3:49:16 PM

Quant Time: Jan 16 22:01:15 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan
QLast Update : Thu Jan 14 16:13:38 2021
Response via : Initial Calibration

