

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL011619\
 Data File : VL033112.D
 Acq On : 17 Jan 2019 4:29
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Jan 17 06:33:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL010419AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jan 04 16:09:05 2019
 QLast Update : Fri Jan 04 16:09:05 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	1.861	2.075	-11.5	101	0.00
3	Chlorodifluoromethane	1.132	1.155	-2.0	103	0.00
4	Chloromethane	0.648	0.655	-1.1	110	0.00
5 T	Vinyl Chloride	0.655	0.667	-1.8	107	0.00
6 T	Bromomethane	0.445	0.435	2.2	104	0.00
7	Chloroethane	0.256	0.252	1.6	108	0.00
8 T	Dichlorotetrafluoroethane	1.603	1.640	-2.3	103	0.00
9 T	Propene	0.423	0.494	-16.8	116	0.00
10 T	Heptane	1.204	1.252	-4.0	103	0.00
11 T	Trichlorofluoromethane	2.056	1.736	15.6	97	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.300	1.198	7.8	97	0.00
13	Ethanol	0.169	0.102	39.6#	91	0.00
14 T	Bromoethene	0.574	0.557	3.0	104	0.00
15 T	Acetone	1.384	1.289	6.9	114	0.00
16 T	1,3-Butadiene	0.522	0.578	-10.7	116	0.00
17	tert-Butyl alcohol	0.293	0.282	3.8	102	0.00
18 T	1,1-Dichloroethene	0.599	0.557	7.0	100	0.00
19 T	Isopropyl Alcohol	0.953	0.682	28.4	93	0.00
20 T	Methylene Chloride	0.577	0.476	17.5	100	0.00
21 T	Allyl Chloride	0.858	0.880	-2.6	102	0.00
22 T	trans-1,2-Dichloroethene	0.594	0.559	5.9	101	0.00
23 T	Vinyl Acetate	1.761	1.871	-6.2	109	0.00
24 T	1,1-Dichloroethane	1.495	1.385	7.4	104	0.00
25 T	Ethyl Acetate	2.340	2.206	5.7	102	0.00
26 T	Hexane	1.102	1.033	6.3	103	0.00
27 T	Carbon Disulfide	1.359	1.397	-2.8	101	0.00
28 T	Methyl tert-Butyl Ether	1.720	1.762	-2.4	108	0.00
29 T	Chloroform	1.877	1.688	10.1	97	0.00
30 T	Cyclohexane	0.821	0.901	-9.7	106	0.00
31 T	cis-1,2-Dichloroethene	1.023	1.080	-5.6	109	0.00
32 T	1,1,1-Trichloroethane	1.905	1.791	6.0	100	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
34 T	2-Butanone	0.581	0.552	5.0	105	0.00
35 T	Carbon Tetrachloride	0.859	0.761	11.4	99	0.00
36 T	Benzene	0.819	0.846	-3.3	109	0.00
37 T	1,2-Dichloroethane	0.572	0.514	10.1	98	0.00
38 T	Trichloroethene	0.337	0.337	0.0	103	0.00
39 T	1,2-Dichloropropane	0.308	0.307	0.3	107	0.00
40 T	1,4-Dioxane	0.117	0.112	4.3	108	0.00
41 T	Tetrahydrofuran	0.296	0.310	-4.7	107	0.00
42 T	Bromodichloromethane	0.786	0.715	9.0	99	0.00
43	Methyl Methacrylate	0.307	0.327	-6.5	104	0.00
44 T	2,2,4-Trimethylpentane	1.387	1.364	1.7	104	0.00
45 T	t-1,3-Dichloropropene	0.383	0.434	-13.3	98	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL011619\
 Data File : VL033112.D
 Acq On : 17 Jan 2019 4:29
 Operator : SY/AP
 Sample : VSTDCCC010
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Jan 17 06:33:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL010419AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Jan 04 16:09:05 2019
 QLast Update : Fri Jan 04 16:09:05 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	cis-1,3-Dichloropropene	0.437	0.511	-16.9	103	0.00
47 T	1,1,2-Trichloroethane	0.360	0.335	6.9	100	0.00
48 T	Dibromochloromethane	0.635	0.613	3.5	96	0.00
49 T	Bromoform	0.555	0.489	11.9	89	0.00
50 T	4-Methyl-2-Pentanone	0.736	0.763	-3.7	102	0.00
51 T	2-Hexanone	0.533	0.574	-7.7	99	0.00
52 T	Tetrachloroethene	0.290	0.293	-1.0	101	0.00
53 T	Toluene	0.878	0.950	-8.2	102	0.00
54 T	1,2-Dibromoethane	0.518	0.515	0.6	98	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
56	1,1,1,2-Tetrachloroethane	0.517	0.433	16.2	94	0.00
57 T	Chlorobenzene	0.835	0.727	12.9	97	0.00
58 T	Ethyl Benzene	1.263	1.272	-0.7	97	0.00
59 T	m/p-Xylene	1.168	1.077	7.8	96	0.02
60 T	o-Xylene	1.143	1.023	10.5	92	0.00
61 T	Styrene	0.730	0.765	-4.8	95	0.00
62	Isopropylbenzene	1.627	1.439	11.6	94	0.00
63 T	1,1,2,2-Tetrachloroethane	0.849	0.673	20.7	90	0.00
64	n-propylbenzene	0.413	0.367	11.1	94	0.00
65	tert-Butylbenzene	1.441	1.264	12.3	93	0.00
66 T	Benzyl Chloride	0.822	0.727	11.6	87	0.00
67	sec-Butylbenzene	2.063	1.789	13.3	92	0.00
68 S	1-Bromo-4-Fluorobenzene	0.767	0.718	6.4	92	0.00
69	p-Isopropyltoluene	1.674	1.510	9.8	95	0.00
70	n-Butylbenzene	1.683	1.553	7.7	98	0.00
71	2-Chlorotoluene	1.154	1.059	8.2	92	0.00
72 T	4-Ethyltoluene	1.257	1.166	7.2	93	0.00
73 T	1,3,5-Trimethylbenzene	1.240	1.111	10.4	93	0.00
74 T	1,2,4-Trimethylbenzene	1.373	1.154	16.0	93	0.00
75 T	1,3-Dichlorobenzene	0.808	0.669	17.2	97	0.00
76 T	1,4-Dichlorobenzene	0.736	0.653	11.3	97	0.00
77 T	1,2-Dichlorobenzene	0.728	0.654	10.2	99	0.00
78 T	Hexachloro-1,3-Butadiene	0.559	0.571	-2.1	150#	0.00
79 T	Naphthalene	1.157	1.370	-18.4	117	-0.01
80 T	Naphthalene,2-methyl-	0.351	0.339	3.4	77	0.00
81 T	1,2,4-Trichlorobenzene	0.559	0.634	-13.4	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0