

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011923\
 Data File : VL040125.D
 Acq On : 19 Jan 2023 22:44
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/20/2023
 Supervised By :Mahesh Dadoda 01/20/2023

Quant Time: Jan 20 02:11:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2023
 QLast Update : Mon Jan 16 14:40:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.208	49	935453	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.674	114	2585760	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.494	117	2278189	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.809	95	1794827	9.479	ppbv	0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.800%
Target Compounds						
					Qvalue	
13) Ethanol	3.269	45	4481m	1.044	ppbv	
20) Methylene Chloride	3.944	84	18700	0.378	ppbv	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL011923\
Data File : VL040125.D
Acq On : 19 Jan 2023 22:44
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/20/2023
Supervised By : Mahesh Dadoda 01/20/2023

Quant Time: Jan 20 02:11:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Jan 16 14:40:47 2023
Response via : Initial Calibration

