

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012021\
 Data File : VL036244.D
 Acq On : 20 Jan 2021 12:21
 Operator : SY/AP
 Sample : VL0120ABL01
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0120ABL01

Manual Integrations
 APPROVED

MMDadoda
 1/21/2021 10:35:55 AM

Quant Time: Jan 21 01:34:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
 QLast Update : Thu Jan 14 16:13:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.677	49	1456456	10.00	ppbv	0.02
33) 1,4-Difluorobenzene	7.182	114	3102175	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.098	117	2762620	10.00	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.426	95	2341139	10.24	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.40%
Target Compounds						
15) Acetone	3.880	43	44592m	0.25	ppbv	Qvalue
20) Methylene Chloride	4.346	84	18475m	0.32	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012021\
Data File : VL036244.D
Acq On : 20 Jan 2021 12:21
Operator : SY/AP
Sample : VL0120ABL01
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0120ABL01

Manual Integrations
APPROVED

MMDadoda
1/21/2021 10:35:55 AM

Quant Time: Jan 21 01:34:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
QLast Update : Thu Jan 14 16:13:38 2021
Response via : Initial Calibration

