

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012021\
 Data File : VL036257.D
 Acq On : 20 Jan 2021 22:02
 Operator : SY/AP
 Sample : QC010421 CAN 10334
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC010421 CAN 10334

Manual Integrations
 APPROVED

MMDadoda
 1/21/2021 10:36:21 AM

Quant Time: Jan 21 01:36:29 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
 QLast Update : Thu Jan 14 16:13:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.665	49	1370920	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.169	114	2957559	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.079	117	2696819	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.413	95	2259658	10.12	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.20%
Target Compounds						
						Qvalue
9) Propene	3.009	41	10641m	0.13	ppbv	
15) Acetone	3.864	43	97300	0.58	ppbv #	76
20) Methylene Chloride	4.337	84	28381m	0.53	ppbv	
26) Hexane	5.684	57	15521m	0.11	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL012021\
Data File : VL036257.D
Acq On : 20 Jan 2021 22:02
Operator : SY/AP
Sample : QC010421 CAN 10334
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC010421 CAN 10334

Manual Integrations
APPROVED

MMDadoda
1/21/2021 10:36:21 AM

Quant Time: Jan 21 01:36:29 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
QLast Update : Thu Jan 14 16:13:38 2021
Response via : Initial Calibration

